



Republic of the Philippines  
Department of Health  
**SAN LAZARO HOSPITAL**

Manila, Philippines  
Telephone Nos.: 8732-3776 to 78; 8732-3106  
E-mail Address: sanlazarohospital@yahoo.com  
Official Website: [www.slh.doh.gov.ph](http://www.slh.doh.gov.ph)



**BIDS AND AWARDS COMMITTEE 4**

**ITB BAC4-2020-002**

**Bid Bulletin No. 001-BAC4-2020**

**TO : INTERESTED BIDDERS**

**DATE : SEPTEMBER 28 , 2020**

**SUBJECT : CLARIFICATIONS/CHANGES ON ITB BAC4-2020-002**

<b>CLARIFICATIONS/CHANGES ON TOR/BOQ FOR THE FOLLOWING CATEGORIES/ITEMS</b>	<b>FROM</b>	<b>TO</b>
A.2 Suction Machine	Suction bottle capacity: 1 or 2 1000 ml	Suction bottle capacity: 2 1L
A.3 Defibrillator with monitor	Rechargeable Lithium Ion Battery	Stick to Hospital specification
A.8 Motorized Treatment Chair	Dimension: 182-186 cm (L) x 62-80 cm (W) x 67-86 cm (H)	182-186 cm (L) x 62-88 cm (W) with arm rest x 67-96 cm (H)
	With pedal/foot control	3 control buttons on the back rest and 4 control buttons on the right side of the bed
	Maximum lift capacity: 385- 441 lbs	Maximum lift capacity: 385-441 lbs, convert kg to lbs and see if your specification is within the set range
		Motorized Treatment Bed can be considered as long as it meets the minimum specifications we have set
A.9 Smoke Evacuator	Suction Power: Max 300W to 400W	Suction Power: As long as it is adjustable, we can consider Max 1200W



Republic of the Philippines  
Department of Health  
**SAN LAZARO HOSPITAL**

Manila, Philippines  
Telephone Nos.: 8732-3776 to 78; 8732-3106  
E-mail Address: sanlazaro-hospital@yahoo.com  
Official Website: [www.slh.doh.gov.ph](http://www.slh.doh.gov.ph)



	Minimum Height requirement: 585 - 740	Stick to Hospital specification
A.10 Bassinet	Mattress with leather covering	Mattress with leatherette covering
	Polycarbonate basket	PVC is good as long as it can hold the patient weight and it is durable
	Overall Dimension: 31"-32" L x 17"-18" W x 37"-39" H	Overall Dimension: 27.95" L x 16.92" W x 33.46" H offer can be considered because it is not too far from the required dimension
A.11 Incubator	Oxygen Consideration: Control Range: 21% to 80%	Oxygen Consideration: Control Range: 21% to 65% offer is good and within the range
	Water Reservoir: 1000 - 1500 ml	Water Reservoir: 1000 ml offer is fine.
	Over all Dimension: 25.1 - 38 x 52.5 - 63.7 x 40 in	Length of 4.1 in is ok as long as the unit is designed to accommodate twin babies
	Hood: 32.6 x 18.8 x 19.6 in	Hood: 33.7 x 18.8 x 19.7 offer is fine
	Mattress: 27.7 - 32" x 0.7 x 14.5 - 16"	Mattress: 25.59" x 13.77" offer is fine
A. 15 Blood Warmer	Temperature Setting: 37-41 °C	Stick to Hospital specification
A.16 Ordinary Laryngoscope	Blades not included in the Hospital Specification	H4 ward uses the following sizes:  MAC2-11 cm (+/-0.5 cm) MAC3 - 13 cm (+/-0.5 cm) MAC 4 - 15 cm (+/-0.5 cm)  Pedia:  Fiber Optic MCINTOSH balde #2-1pc/#3-1 pc/#4-1pc Fiber Optic Miller #001-1 pc/#0- 1 pc/#1-1pc
	Battery (Rechargeable/Not Rechargeable) is not Specified in the Hospital Specification	Stick to Hospital specification



Republic of the Philippines  
 Department of Health  
**SAN LAZARO HOSPITAL**

Manila, Philippines  
 Telephone Nos.: 8732-3776 to 78; 8732-3106  
 E-mail Address: sanlazaro@hospital@yahoo.com  
 Official Website: [www.slh.doh.gov.ph](http://www.slh.doh.gov.ph)




A. 17 Infusion pump	Safety Features: Able to warm IV Solution from 25OC to 45OC over temperature protection	Safety Features: Able to warm IV Solution from 25° C to 45° C over temperature protection
	Power Supply: 25VA	Power Supply: 35VA is acceptable as long as the unit will function
	Bolus:0.1 ml/h up to speed 1500 ml/h	Stick to Hospital specification
	High Precision Infusion (ACCURACY) : +/-3%	Stick to Hospital specification
	Battery: Rechargeable Lithium battery, at least 8 hours working time after fully charged	Stick to Hospital specification
A. 21 Walk -in freezer	Specification for Compressor not indicated	Semi-Hermetic Compressor
		The compression strength is minimum at 150 kpa or more, people can walk on the floor panel and allow trolleys to go in and out
		With plastic curtain in front of the door freezer
B.2 Mini Pharmaceutical Refrigerator	Auto defrost removed from Hospital Specification	Include Auto defrost

Please be guided accordingly.

Prepared By:

  
**FAYE MICHELLE P. MENDOZA, RND**  
**BAC Secretariat**

Approved By:

  
**DAVID T. SUPLIGO, MD**  
**Chairperson-Bids and Awards Committee 4**