



Republic of the Philippines  
Department of Health  
**SAN LAZARO HOSPITAL**

Manila, Philippines  
Telephone Nos.: 8732-3776 to 78; 8732-3106  
E-mail Address: sanlazarohospital@yahoo.com  
Official Website: [www.slh.doh.gov.ph](http://www.slh.doh.gov.ph)



**BIDS AND AWARDS COMMITTEE 4**

**ITB BAC4-2020-002**

**Bid Bulletin No. 003**

**TO : INTERESTED BIDDERS**

**DATE : OCTOBER 5, 2020**

**SUBJECT : CLARIFICATIONS/CHANGES ON ITB BAC4-2020-002**

| <b>CLARIFICATIONS/CHANGES ON TOR/BOQ FOR THE FOLLOWING CATEGORIES/ITEMS</b> | <b>FROM</b>  | <b>TO</b>  |
|---|--|--|
| A.2 Suction Machine   | Suction bottle capacity:<br>1 or 2 1000 ml   | Suction bottle capacity:<br>2 1L   |
| A.8 Motorized Treatment Chair   | Dimension: 182-186 cm (L)<br>x 62-80 cm (W) x 67-86 cm (H)   | 182-186 cm (L) x 62-88 cm (W) with arm rest x 67-96 cm (H)                               |
|   | 3 control buttons on the back rest and 4 control buttons on the right side of the bed                | with pedal/foot control and/or control buttons on back rest and/or on either side of bed |
|   | Motorized Treatment Bed can be considered as long as it meets the minimum specifications we have set | Motorized Treatment Bed/Chair  |
| A.9 Smoke Evacuator   | Suction Power:<br>Max 300W to 400W   | Suction Power:<br>As long as it is adjustable, we can consider Max 1200W                 |
| A.10 Bassinet   | Mattress with leather covering   | Mattress with leatherette covering   |
|   | PVC is good as long as it can hold the patient weight and it is durable                              | Polycarbonate basket/PVC   |



Republic of the Philippines  
Department of Health  
**SAN LAZARO HOSPITAL**

Manila, Philippines  
Telephone Nos.: 8732-3776 to 78; 8732-3106  
E-mail Address: sanlazarohospital@yahoo.com  
Official Website: [www.slh.doh.gov.ph](http://www.slh.doh.gov.ph)




|                            |  |  |
|----------------------------|--|--|
|                            | Overall Dimension:<br>27.95" L x 16.92" W x 33.46" H<br>offer can be considered because it is not too far from the required dimension  | Overall Dimension:<br>25-32" (L) x 15-18" (W) x 30-39" (H)   |
| A.11 Incubator             | Oxygen Consideration:<br>Control Range: 21% to 65%<br>offer is good and within the range   | Oxygen Concentration:<br>Control Range: 21% to maximum of at least 65%                                   |
|                            | Length of 4.1 in is ok as long as the unit is designed to accommodate twin babies  | Over all Dimension: 25.1 – 38" x 52.5 - 63.7" x 40-42"   |
|                            | Hood: 33.7 x 18.8 x 19.7 offer is fine   | Hood: 32-34" x 18-20" x 19-21"   |
|                            | Mattress: 25.59" x 13.77" offer is fine  | Mattress: 25-32" x 0.7-1" x 13-16"   |
| A.16 Ordinary Laryngoscope | H4 ward uses the following sizes:<br><br>MAC2-11 cm (+/-0.5 cm)<br>MAC3 – 13 cm (+/-0.5 cm)<br>MAC 4 – 15 cm (+/-0.5 cm)<br><br>Pedia:<br><br>Fiber Optic MCINTOSH balde<br>#2-1pc/#3-1 pc/#4-1pc<br>Fiber Optic Miller<br>#001-1 pc/#0- 1 pc/#1-1pc | 3 blades per sets, winning supplier will communicate with agency with regards to size and type of blades |
|                            | 2 sets (TOR)   | 8 sets (TOR)   |

Please be guided accordingly.

Prepared By:

  
**TERESA A. GUERRERO, OTRP**  
**BAC4 Secretariat**

Approved By:

  
**DAVID T. SUPLICO, MD**  
**Chairperson-Bids and Awards Committee 4**



Republic of the Philippines  
Department of Health  
**SAN LAZARO HOSPITAL**

Manila, Philippines  
Telephone Nos.: 8732-3776 to 78; 8732-3106  
E-mail Address: sanlazarahospital@yahoo.com  
Official Website: [www.slh.doh.gov.ph](http://www.slh.doh.gov.ph)



**NEW SCHEDULE FOR BAC ACTIVITIES**

| <b>BAC ACTIVITIES</b>   | <b>SCHEDULE DATE/TIME</b>   | <b>VENUE</b>     |
|---|-----------------------------|------------------|
| Submission and Opening of Bid Proposals for<br><br><b>A. Procurement of Medical and Dental Equipment for 2020</b> |                             |                  |
| Item Nos. 2,8,9,10,11,16  | October 15, 2020<br>9:00 AM | SLH Amphitheater |

**DAVID T. SUPLICO, MD**  
Chairperson-Bids and Awards Committee 4