

SLH BIDS AND AWARDS COMMITTEE FOUR
BAC Resolution No. PB-021 s. 2020

PROCUREMENT OF GOODS: MEDICAL AND DENTAL EQUIPMENT FOR 2020 (Category A.3: DEFIBRILLATOR WITH MONITOR)

WHEREAS, the San Lazaro Hospital, through its Bids and Awards Committee Four, posted an Invitation to Bid No. BAC4-2020-002: Procurement of Goods - Medical and Dental Equipment on September 11, 2020 with fund source from GAA2020/HFEP at PhilGEPS website, SLH website and in other conspicuous places in the hospital;

WHEREAS, on September 21, 2020, the following bidder attended the Pre-Bid Conference. Subsequently, the following bidder also procured the bidding documents:

Category No.	Description	Suppliers (Attended the Pre Bid Conference)	Suppliers (Procured Bidding Documents)
A.3	DEFIBRILLATOR WITH MONITOR	MEDEV Medical Devices Corp	MEDEV Medical Devices Corp

WHEREAS, on September 28, 2020, Bid Bulletin No. 001-BAC4-2020 was issued for this category:

Cat.	Description	Changes From (Old Specification)	To (New Specification)
A.3	DEFIBRILLATOR WITH MONITOR	Schedule of Bid Opening: October 1, 2020, 9:00 AM	New Schedule of Bid Opening: October 6, 2020 9:00 AM

WHEREAS, only **MEDEV Medical Devices Corp.** participated for the abovementioned category during the conducted Submission and Opening of Bids last October 6, 2020;

WHEREAS, **MEDEV Medical Devices Corp.** passed the Preliminary Examination of Bids. Envelope B was opened and their offered bids are follows:

Category A.3: Defibrillator with Monitor				ABC		Offered Bid	
Item No.	Quantity and Unit of Measure	Hospital Specifications	Unit Price	Total Price	MEDEV Medical Devices Corp.	Remarks	
1	1 unit	Compact, Light weight with adult and pedia paddle Biphasic defibrillator monitor Able to deliver shocks from 1 Joule to 200 Joules Should monitor ECG through paddles, pads and monitoring electrodes Shock delivery through pads and paddles Body impedance for a range of 15 to 300 ohms Built in thermal paper printer Charging time of less than 7 seconds for maximum energy LED/LCD 6.5- 8.4 inch Colour display for viewing monitoring parameters like ECG, SpO2, NIBP and etCO2 Power Supply: Mains: 220 V/ 60 Hz Rechargeable Lithium Ion battery capable of usage for at least 5 hours of monitoring	PHP 300,000.00	PHP 300,000.00	PHP 299,000.00 (Brand Offered: Comen S6)	SCB	

Handwritten signature and stamp.



SAN LAZARO HOSPITAL

Manila, Philippines

Telephone Nos.: 8-732-3776 to 78; 8-732-3106

E-Mail Address: slhbac4@gmail.com



WHEREAS, on October 13, 2020, Notice of Post Qualification was sent to the bidder;


WHEREAS, on October 15, 2020, post qualification and product demonstration were done by the TWG and End User; and the offered bid of **MEDEV Medical Devices Corp.** for Category A.3 was found to be complying;


WHEREAS, on October 19, 2020, BAC4, together with the TWG and End User, deliberated on the findings and the offered bid of **MEDEV Medical Devices Corp.** for Category A.3 was found to be responsive;

NOW THEREFORE, We, the members of the **Bids and Awards Committee Four**, hereby RESOLVE as it is RESOLVED:

1. To declare **MEDEV Medical Devices Corp.** as the **Single Calculated and Responsive Bid (SCRB)** for **Category A.3. DEFIBRILLATOR WITH MONITOR**;
2. To recommend the award of contract to **MEDEV Medical Devices Corp.** with a total contract price of **Two Hundred Ninety Nine Thousand Pesos (PHP 299,000.00)**;
3. To submit this recommendation to the Medical Center Chief II for approval.


RESOLVED at San Lazaro Hospital, Manila, this 20th day of October 2020.


DAVID T. SUPLICO, MD
 BAC4 Chairperson


BRENDA DC. TAGUBA, RSW, RN, MGM
 Member


DANIEL S. TAGULAO, RND
 Vice-Chairperson


NIKKO LAE D. ABDON, RN, MAN
 Member


JESUS MARIO L. LOGMAO, MD
 Provisional Member

Approved as Recommended:


EDMUNDO B. LOREZ, MD, MPH, MHA
 Medical Center Chief II

