



**SLH BIDS AND AWARDS COMMITTEE FOUR
BAC Resolution No. PB-041 s. 2020**

**PROCUREMENT OF GOODS: MEDICAL AND DENTAL EQUIPMENT FOR 2020
(Category A.7: INFUSION PUMP)**

WHEREAS, the San Lázaro Hospital, through its Bids and Awards Committee Four, posted an Invitation to Bid No. BAC4-2020-004 : Procurement of Goods - Medical and Dental Equipment on November 19, 2020 with fund source from GAA2020/HFEP at PhilGEPS website, SLH website and in other conspicuous places in the hospital;

WHEREAS, on November 27, 2020, the following bidders attended the Pre- Bid Conference. Subsequently, the following bidders also procured the bidding documents:

Category No.	Description	Suppliers (Attended the Pre Bid Conference)	Suppliers (Procured Bidding Documents)
A.7	Infusion Pump	MEDEV Medical Devices Corp., Asya Medika Inc., Medicotek Inc., Enhanced Medical and Industrial Ent., Variance Trading Corp., Dynamed Healthcare Inc., Endure, Haggai Medical and Blue Sky trading Inc.	Asya Medika Inc., Enhanced Medical and Industrial Ent., Dynamed Healthcare Inc. MEDEV Medical Devices Corp.

WHEREAS, on December 1, 2020, Bid Bulletin No. 006 BAC4-2020-004 was issued for this category:

Cat.	Description	Changes From (Old Specification)	To (New Specification)
A.7	Infusion Pump	Functions: 1. Bolus : 0.1 ml/h up to speed 1500ml/h	Functions: 1. Bolus : 0.1 ml/h up to speed of 1000ml/h
		2. Accuracy : < + / - 3 %	< + / - 5 %
		Power Supply, 60 Hz, 25 VA	25 VA or 35 VA
		Battery : at least 8 hours working time	At least 6 hours working hours
		Safety Features: 250C to 450C	25° to 45°
		Max Weight: 2 kg	2.5 kg
		One (1) infusion Pump Unit should come with: 3. 10 pieces IV sets adult 4. 10 pieces IV sets pedia	3.10 boxes IV sets adult 4. 10 boxes IV sets pedia

WHEREAS, on December 9, 2020, Four (4) suppliers namely: (1) **Asya Medika Inc.**, (2) **Dynamed Healthcare Inc.**, (3) **MEDEV Medical Devices Corp.**, (4) **Enhanced Medical and Industrial Ent.** participated for the above mentioned category during the conducted Submission and Opening of Bids;

WHEREAS, all bidders passed the Preliminary Examination of Bids. Envelope B were opened and their offered bids are as follows:



Category A.7: Infusion Pump									
Item No.	Quantity and Unit of Measure	Hospital Specifications	ABC		Offered Bid				
			Unit Price	Total Price	(1) Asya Medika Inc.	(2) Dynamed Healthcare Inc.	(3) MEDEV Medical Devices Cor.	(4) Enhanced Medical and Industrial Ent.	Remark
A.7	30units	<p>Functions:</p> <ol style="list-style-type: none"> Bolus: 0.1m/h up to speed1500ml/h Flow rate range: 1-600ml/h Volume limit: 1-9999 ml High Precision Infusion (ACCURACY)+/- 3% <p>Display Monitor At least 4 inches LCD Display</p> <p>Power supply: AC 220V, 60 Hz, 25VA Battery: Rechargeable lithium battery, at least 8 hours working time after fully charge Electrical safety: compliance with the requirements of IEC 60601-1;waterproof design (avoids water penetration)</p> <p>Safety Features: *Automatic detect door closing situation can effectively prevent drug flowing and protect the tube * Able to warm IV solution from 25°C to 45°C over temperature protection * Software adjusting function suit for any brand of infusion administration sets * Universal pump that could cater different brands of IV set * At least 800 infusion records and last configuration memory * With anti-bolus system to reduce significant bolus after occlusion sudden release * Precise pressure monitoring sensor - able to detect down to 50mmhg with 1ms, alarm will be triggered once blocking happens</p> <p>Physical Specification Manufacturer specs allowed max. weight: 2kg One (1) Infusion Pump Unit should come with:</p> <ol style="list-style-type: none"> 1 piece bracket 1 piece IV pole with atleast 4 wheels and atleast 2 hooks 1 piece drop sensor 10 pieces IV sets Adult 10 pieces IV sets Pedia <p>EMC: Complies with IEC/EN60601-1-2 and IEC/EN60601-2-24.</p>	120,000.00	3,600,000.00	2,558,640.00	2,249,940.00	2,246,640.00	1,800,000.00	LCB

WHEREAS, on December 9, 2020, Notice of Post Qualification was sent to the Lowest Calculated Bidder;




WHEREAS, on December 16, 2020, post qualification and product demonstration were done by the TWG and End User; and the offered bid of **Enhanced Medical and Industrial Enterprises** for Category A.7 was found to be complying;

WHEREAS, on December 17, 2020, BAC4, together with the TWG and End User, deliberated on the findings and the offered bid of for Category A.7 was found to be responsive;

NOW THEREFORE, We, the members of the **Bids and Awards Committee Four**, hereby RESOLVE as it is RESOLVED:

1. To declare **Enhanced Medical and Industrial Enterprises** as the **Lowest Calculated and Responsive Bid (LCRB)** for **Category A.7. Infusion Pump**;
2. To recommend the award of contract to **Enhanced Medical and Industrial Enterprises** with a total contract price of **One Million Eight Hundred Thousand Pesos (PHP 1,800,000.00)**;
3. To submit this recommendation to the Medical Center Chief II for approval.

RESOLVED at San Lazaro Hospital, Manila this 18th day of December 2020.


DAVID T. SUBLICO, MD
BAC4 Chairperson



BRENDA DC. TAGUBA, RSW, RN, MGM
Member


DANIEL S. TAGULAO, RND
Vice-Chairperson


NIKKO LAE D. ABDON, RN, MAN
Member

JESUS MARIO L. LOGMAO, MD
Provisional Member

Approved as Recommended:


EDMUNDO B. LOPEZ, MD, MPH, MHA
Medical Center Chief II



December 18, 2020

NOTICE OF AWARD

ENHANCED MEDICAL AND INDUSTRIAL ENTERPRISES


2ND Floor bldg. Golden Matina Complex
97 McArthur Highway, Davao City
Tel No. 082-222-1077
E-mail Add: enhanced_mie11@yahoo.com

Dear Sir/Madam;

This is to inform you that based on the results of Public Bidding on December 9, 2020 for the **Procurement of Goods: Medical and Dental Equipment** with a total contract price of **One Million Eight Hundred Thousand Pesos (Php 1,800,000.00)** inclusive of local taxes, your proposal for **Categories A.7** is hereby accepted, *viz.*

Category A.7 : Infusion Pump			Total No. of Quantity	Unit	Offered Bid		Remarks
Item No.	Particulars	Specifications			Unit Price	Total Price	
1	Infusion Pump	Fast injection/Bolus: 1500ml/h, Flow rate range: 0.1-1500ml/h, Volume limit: 1-9999 ml, High Precision Infusion (ACCURACY): ±3%, 5.4 inches LCD Screen, 220V, 50/60Hz, Rechargeable lithium ion battery 2200mAh: 8 hours continuous working @25ml/h, ISO 13485:2003:IPX4 water proof level, Automatic test the closeness of pump door: prevent drug flow by gravity in infusion tube, IV warming with over temperature protection, Customizable/ universal IV brands, Universal pump that could cater different brands of IV set, over 800 infusion records, Discrete bubble sensor for anti-bolus system, High sensitivity for precise monitoring, 1.8kg, 1 piece Bracket, 1 piece IV pole with at least 4 wheels and at least 2 hooks, 1 piece drop sensor, 1 box IV sets Adult or 100 pieces, 1 box IV sets Pedia or 100 pieces, EMC: Complies with IEC/EN60601-1-2 and IEC/EN60601-2-24.	30	Units	PHP 60,000.00	PHP 1,800,000.00	LCRB
TOTAL AMOUNT						PHP 1,800,000.00	

Relative thereto, you are hereby directed to submit, within ten (10) calendar days from receipt of this NOA your **PERFORMANCE SECURITY** in any form stipulated in **Section 39 of 2016 Revised IRR of Republic Act 9184.**





Failure to provide the warranty security shall constitute sufficient ground for cancellation of the award and forfeiture of the bid security.

Please be guided accordingly.

Very truly yours,


EDMUNDO B. LOPEZ, MD, MPH, MHA
Medical Center Chief II

CONFORME:

Name of Company

Name and Signature of Authorized Signatory & Date