



Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazaro@hospital@yahoo.com
Official Website: www.slh.doh.gov.ph



BIDS AND AWARDS COMMITTEE 2

Bid Bulletin No. 01
ITB BAC2-2020-003

TO: ALL CONCERNED SUPPLIERS/BIDDERS
DATE: November 20, 2020
SUBJECT: Clarifications/Changes for all **CATEGORIES** on ITB BAC2-2020-003:

Changes to the Date of submission/opening of bids from **ALL** Categories.

Pursuant to the mandate of Revised IRR of RA 9184 under section 22.5.2 the advertised opening of bids under ITB BAC2-2020-003 shall be moved as follows:

ALL CATEGORIES

Opening of Bids (Previous date)	Opening of Bids (Latest date)
November 24, 2020, 9:30 AM	December 1, 2020, 9:30 AM

This pertains to the queries of the prospected bidders in reference to Pre-Bid conference inquiries dated November 6, 2020, 9:30 AM at the SLH Amphitheater, with Invitation to Bid number (BAC2-2020-003).

Queries	Answers
1. Biomedical Freezer, -20°C to 40°C	
<p>Zafire Distributors, Inc.</p> <p>"Offer 121L for Ref, 120L for Freezer 4 rollers that is lockable to secure Temp. -20 to -32°C 2 Drawers and 2 wire shelves."</p>	<p>"We have to follow our Biomed Freezer – 20-40°C with Refrigerator specs as described on the PR." - Dr. Zita Dela Merced</p>
<p>Omnibus Bio-Medical Systems, Inc.</p> <p>Two-door biomedical upright freezer (doors with locks); Working temperature range for Refrigerator: 2-8°C and for Freezer: -20 to -30°C: Capacity at least 80-98 L; Exterior: Painted steel; Interior: painted aluminum plate Defrosting system: Manual defrost: Refrigerant: CFC free; Shelves (Max Load): 3 shelves Ref 1 shelf freezer</p>	<p>"We have to follow our Biomedical Freezer -20 to -40°C with Refrigerator specifications that includes 3 shelves for the Refrigerator and 2 shelves for the freezer" - Dr. Zita Dela Merced</p>
<p>Guill-Bern</p> <p>Sprayed steel to be considered for interior and exterior paint</p>	<p>"Yes, it can be accepted". - Dr. Zita Dela Merced</p>





Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazarahospital@yahoo.com
Official Website: www.slh.doh.gov.ph



<p>Can 4 casters (2 front caster adjustable and 2 back casters fixed) be considered?</p>	<p>We have to follow our specifications for the said item/ equipment as described - Dr. Zita Dela Merced</p>
<p>2. Biomedical/Biological Refrigerator, 13-14 cu. ft.</p>	
<p>Omnibus Bio-Medical Systems, Inc.</p> <p>Single door biomedical upright refrigerator (12 -14 cu. ft. with working temp, range: 2 -8°C): CFC free cooling system; Clear transparent , lockable door Refrigerator compartment showcase type (at least 340-390L capacity) with at least 4-6 easily cleaned, adjustable shelves;Automatic defrost; Microprocessor control, Independent/separate cooling system; With caster wheels; Digital LED display temperature display; with remote alarm system ready with Internal lamp/light</p> <p>Guill-Bern</p> <ul style="list-style-type: none"> • 23.16 cubic feet, can be considered? • 10 adjustable shelves, can this be considered? • Double machine cascade refrigeration for the cooling system, can be considered? • Audible and audio alarm, can be considered if remote alarm system is not available 	<p>"Regarding the Biomedical/Biological Refrigerator, we will accept if you can provide a 14 cu. Ft with at least 390L - Dr. Zita Dela Merced</p> <p>We have to follow our specifications for the said item/ equipment as described - Dr. Zita Dela Merced</p>
<p>3. Microcentrifuge, 30 placer</p>	
<p>Rainphil, Inc.</p> <p>For Microcentrifuge, can we give a minimum at least 24-30 placer?</p> <p>Grepcor Diamonde, Inc</p> <p>For Microcentrifuge, can we give a minimum at least 24-30 placer?"</p>	<p>"We need at least 30 placer Microcentrifuge" - Dr. Zita Dela Merced</p> <p>"We need at least 30 placer Microcentrifuge" - Dr. Zita Dela Merced</p>
<p>4. Dry Bath</p>	
<p>Rainphil, Inc.</p> <p>For Heating Block, can we give at least 36-40 holes?</p>	<p>For heating block, we can accept the 36-40 holes - Dr. Zita Dela Merced</p>
<p>5. Nucleic Acid Extraction Machine</p>	
<p>Omnibus Bio-Medical Systems, Inc.</p> <p>Isolates DNA, RNA and proteins from a variety of stating materials including cell-free body fluids, blood, bacteria, cell cultures, tissue and plant samples; with volumes ranging up to 5 mL; Automated magnetic bead or One tube-technology, allows excellent recovery of samples with great reproducibility and efficiency; compatible with</p>	<p>"We have to follow our own Nucleic Acid Extraction Machine specifications" - Dr. Zita Dela Merced</p>



Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazarohospital@yahoo.com
Official Website: www.slh.doh.gov.ph



<p>existing PCR machine of the end-user or with most common PCR machine</p> <p>Rainphil, Inc.</p> <p>For US connection, is it possible to remove this requirement? Because for automated extraction machine, there is no result to be transferred to the US, We could only see protocols in the machine. We have USB for protocol transport to computer.</p> <p>Volume range of 5ml, for automatic magnetic bad extraction, maximum volume per well is only 2 ml?</p> <p>96 to 100 samples, there is only 32, 47 and 96 placer for automated extraction. Will it be acceptable if we will offer 96-placer?</p> <p>Guill-Bern</p> <ul style="list-style-type: none"> • Can sample volume can be ranged from 1 ml to 5 ml? • Can sample capacity be ranged from 16-100? 	<p>Yes, we can remove the US connectivity, provided you have to provide USB. - Dr. Zita Dela Merced</p> <p>We need 200ul volume for automatic magnetic extraction. - Dr. Zita Dela Merced</p> <p>Yes, we can accept the 96 placer only - Dr. Zita Dela Merced</p> <p>We have to follow our specifications for the said item/ equipment as described - Dr. Zita Dela Merced</p>
<p>6. Biosafety Cabinet Class II with UPS</p>	
<p>Omnibus Bio-Medical Systems, Inc.</p> <p>Type A2 (if NSF49-certified) or Class II (if EN12469-certified), ductless, includes UPS (5 mins. Back up time) and built-in AVR with power delay</p> <p>Nominal size: 4 ft (1.2 m) usable work area: 0.72 m2 (7.8 sq. ft.) ergonomic design</p> <p>Guill-Bern</p> <ul style="list-style-type: none"> • 230V, can this be considered? • Usable work area: (0.81m2) (8.7 sq ft), can this be considered? • Hepa filter 0.3 um 99.97% efficient filter, can this be considered? 	<p>"We have to follow our BSC Class II specifications" - Dr. Zita Dela Merced</p> <p>We have to follow our specifications for the said item/ equipment as described - Dr. Zita Dela Merced</p>
<p>11. Video Laryngoscope</p>	
<p>RGBM Medical Express Sales, Inc.</p> <p>RBGM's offers different variant of sizes in terms of laryngoscope blades.</p> <ul style="list-style-type: none"> • Specific sizes for adult • Specific sizes for pediatric • Specific size of straight blade <p>RBGM's would like to request reconsideration to at least DOH Letters made of sticker on the visible</p>	<p>Pedia Blades</p> <ol style="list-style-type: none"> Miller Blade # 1 (1 unit) Miller Blade # 0 (1 unit) Difficult Blade, Pedia (1 unit) Macintosh Blade # 3 (1 unit) <p>Also, the PIDTMD Department agreed to have the DOH letters be made of water proof stickers instead</p>



Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazaro@hospital@yahoo.com
Official Website: www.slh.doh.gov.ph



part of the equipment.

The embedded letters on the medical equipment may cause corrosion or damage that not recommended by the manufacturer.

Health Solutions (For Pedia)

- Grip handle: Must be of the classic pistol grip type
We respectfully ask for reconsideration of making the specs to be **Manufacturer's specs or low-profile design and ergonomic handle**. Low-profile design improves maneuverability and working space and ergonomic handle design added control and comfort.

- Blades:

One straight blade size

We respectfully ask for reconsideration of making the specs to be **Manufacturer's Specification or Lo Pro T2** as our Lo Pro type of blade has 60-degree angulation which follows the human anatomy. **Blade designed for only 2-3 lbs of lift force (vs 10-12 lbs in Direct Laryngoscopy)** to be easy on patient and low stress on the airway. Low-profile designs are lightweight and streamlined, offering improved maneuverability and working space for routine and difficult intubations. **Recommended patient weight/size is 1.8-10 kg (4-22 lbs).**

- Video screen monitor

- o Must be at least 7 inch size

We respectfully ask for reconsideration of making the specs to be **Manufacturer's Specification or the screen size should be atleast 6.4 inches** as that is already big enough to see the glottic view. Also, our Glidescope Video Monitor has a high-quality video output to an external monitor for OR theatre viewing.

- o 1920x1080 pixel resolution

We respectfully ask for reconsideration of making the specs to be **Manufacturer's Specification or the screen size should be atleast 640 x 480** as that is

of having them embossed on the visible part of the video laryngoscope

- **Dr. Edna Miranda**

Adult Blades

e) Macintosh Blade # 4 (1 unit)

f) Macintosh Blade # 3 (1 unit)

g) Macintosh Blade # 2 (1 unit)

h) Difficulty Blade, Adult (1 unit)

- **Dr. Ann Celestyn P. Uichanco**

Specs as is.

- **Dr. Edna Miranda**

Specs as is.

- **Dr. Edna Miranda**

Specs as is.

- **Dr. Edna Miranda**

Specs as is.

- **Dr. Edna Miranda**



Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazaro@hospital@yahoo.com
Official Website: www.slh.doh.gov.ph



<p>already clear enough to see the Glottic view of the patient.</p>	
<ul style="list-style-type: none">○ With video quality adjustments and multimedia interface	<p>Specs as is.</p>
<p>We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification or high-resolution, full-color digital camera and monitor for real-time viewing and recording, record and external video button. Having those specifications are essential especially when reviewing cases or when conducting a training.</p>	<p>- Dr. Edna Miranda</p>
<p>Health Solutions (For Adult)</p>	
<p>- System Consisting of: Macintosh Blade #2 (1 unit)</p>	<p>Specs as is.</p>
<p>We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification or Lo Pro T2 as our Lo Pro type of blade has 60-degree angulation which follows the human anatomy. Blade designed for only 2-3 lbs of lift force (vs 10-12 lbs in Direct Laryngoscopy) to be easy on patient and low stress on the airway. Low-profile designs are lightweight and streamlined, offering improved maneuverability and working space for routine and difficult intubations.</p>	<p>- Dr. Ann Celestyn P. Uichanco</p>
<p>- Technical Specifications: Monitor for CMOS Endoscopes:</p>	
<ul style="list-style-type: none">❖ Screen resolution 1280 x 800 pixels <p>We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification or the screen size should be atleast 640 x 480 as that is already clear enough to see the Glottic view of the patient.</p>	<p>Specs as is.</p> <p>- Dr. Ann Celestyn P. Uichanco</p>
<ul style="list-style-type: none">❖ Screen size at least 7 inches	<p>Specs as is.</p>
<p>We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification or the screen size should be atleast 6.4 inches as that is already big enough to see the glottic view. Also, our Glidescope Video Monitor has a high-quality video output to an external monitor for OR theatre viewing.</p>	<p>- Dr. Ann Celestyn P. Uichanco</p>



Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazarohospital@yahoo.com
Official Website: www.slh.doh.gov.ph



<p>❖ Two (2) camera output with USB and HDMI port We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification or one camera output as it is already enough to see the glottic view.</p> <p>❖ Atleast 16GB SD card, can capture video and image in real time and data transfer from SD card to USB Flash drive We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification or making the SD card optional. The capture image and recorded case can be transferred directly to USB flash drive which can save time and effort</p> <p>❖ Shock resistant ABS plastic housing We respectfully ask for reconsideration of making the specs to be Manufacturer's Specification. Our Glidescope Video Monitor is protected against dust ingress sufficient to prevent the product from operating normally. The product is fully protected against solid objects and splashing of water from any angle.</p> <p>- Accessories included: ❖ Protection cap We respectfully ask for reconsideration of making this accessory as optional since our titanium blades do not need protection cap which will cost you less. Our Titanium blades are IPX8 compliant, with disinfection and sterilization options. No soaking cap required.</p> <p>CMOS technology Blades: ❖ Reusable Macintosh Bladesize 4, autoclave ❖ Reusable Macintosh Bladesize 3, autoclave ❖ Reusable Macintosh Bladesize 2, autoclave We respectfully ask for reconsideration of making the specs to be Manufacturer's specs. Our titanium blades do not need to autoclave. Titanium Blades are completely submersible in the recommended cleaning solutions</p> <p>❖ Reusable Blade for difficulty intubation, adult with catheter introduction size 16-18 We respectfully ask for reconsideration of making</p>	<p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p> <p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p> <p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p> <p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p> <p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p> <p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p> <p>Specs as is. - Dr. Ann Celestyn P. Uichanco</p>
---	---





Republic of the Philippines
Department of Health
SAN LAZARO HOSPITAL

Manila, Philippines
Telephone Nos.: 8732-3776 to 78; 8732-3106
E-mail Address: sanlazarohospital@yahoo.com
Official Website: www.slh.doh.gov.ph



the specs to be Manufacturer's specs or optional . Our titanium blades are the thinnest blades in the market offering improved maneuverability and working space for routine and difficult intubations.	
SLCC	
Gentleprince, Inc. Since our company just started the venture on joining public bidding last 2018. I would like to ask if we can join the public bidding using private projects amounting 50% of the total ABC of the project?	Yes. - Dr. Efren Dimaano
11. Video Laryngoscope	
From Category 11	11.1 Video Laryngoscope, Pedia 11.2 Video Laryngoscope, Adult
12-15 Air Conditioner	
To include free installation	


EFREN M. DIMAANO, MD
Chairperson, BAC2 (old)

