



**BIDS AND AWARDS COMMITTEE 4**

**ITB BAC2-2020-004**

**Bid Bulletin No. 001**

**TO : INTERESTED BIDDERS**

**DATE : DECEMBER 10, 2020**

**SUBJECT : CLARIFICATIONS/CHANGES ON ITB BAC2-2020-004**

<b>CLARIFICATIONS/CHANGES ON TOR/BOQ FOR THE FOLLOWING CATEGORIES/ITEMS</b>	<b>FROM</b>	<b>TO</b>
1. Biomedical Freezer, -10 to -30 C	b. -10 to -30°C	-10 to -30 °C or -10 to -40°C
	c. 250 L	minimum of 268L
	d. painted steel	painted steel or sprayed steel
	e. painted steel	painted steel or sprayed steel
	h. HFC	HFC or R290
	i. 4-5 adjustable shelves	3- 5 adjustable shelves
	k. 4 casters with adjustable feet	adjustable caster or 4- fixed caster
2. Biological Refriferator	a. 9 cu. Ft, slim type	8-9 cu. Ft, slim type
5. Centrifuge	f. Fixed angle 35 degree rotor g. Adapter for Four-Place, swing-Out Rotor	Swing- out rotor; adapter for Four - Place
6. Vortex Mixer	f. Includes attachments of various shapes, sizes, and materials allows for mixing of almost all common tubes or containers for Hands- free or touch operation	General attachments for all kinds of tubes
8. Minifuge	g. Metal base	Heavy plastic material OR metal base
10. Micropipetors	b. Antimicrobial	Fully autoclavable or antimicrobial
ALL CATEGORIES	Procurement of Infrastructure	Procurement of Goods
	Opening and submission – December 17, 2020 @ 1:30pm	December 17, 2020 @ 9:00 AM

Please be guided accordingly.

Prepared By:

**BRENDA D. TAGUBA**  
Member, Bids and Awards Committee 4

Approved by:

**DAVID T. SUPILICO, MD**  
Chairperson, Bids and Awards Committee 4

